


STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Township Police Department 2400 Bethel Avenue Pennsauken, NJ 08109	2. Contact: Sgt. Chris Sulzbach 3. Phone Number: 856-488-0080	4. Date: 03/04/2015 5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUM-0051</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Complaint received that the Alcotest was experiencing Control Test Failures.



10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed control test failures due to low electrochemical sensor (EC) results. Instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 02303

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Michael P. Gibson #6353
Name & Badge Number (Print)

Michael P. Gibson #6353 3/4/2015
Signature Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Pennsauken Twp. P.D., Pannsauken, NJ

Date Received: 03/13/15

Date given to service: 03/13/15

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0051

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFJ-0630
 CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/17/15

Packing Slip

Customer no Order No Order date
 150056355 10764988 03/17/2015

Packing slip no. **Ship date**
 80932163 03/17/2015

Please reference on inquiries

Ship to
 150049890
 PENNSAUKEN TOWNSHIP
 MUNICIPAL COMPLEX
 2400 BETHEL AVENUE
 PENNSAUKEN, NJ 08109
 USA

Payer

Your Purchase Order Number
 03/17/2015
 15-00611
 Any questions? Please contact:
 ROSEMARIE HUGHES

Customer
 PENNSAUKEN POLICE DEPARTMENT
 ATTN: CHRIS SULZBACH
 CAMDEN COUNTY
 2400 BETHEL AVENUE
 PENNSAUKEN, NJ 08109-2792
 USA

Your sales person
 LINDA SALO
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 2 Complete delivery

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARUM-0051 REC'D 3/13/15			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 1 / 0 EA			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.5 / 0.0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906

Packing Slip

Customer no Order No Order date
150056355 10764988 03/17/2015

Packing slip no. Ship date
80932163 03/17/2015

Please reference on inquiries

Page 2 / 2 Complete delivery

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 1 / 0 EA			
0040	2 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 2 / 0 EA			
0050	1 EA	MPINS INSURANCE AND HANDLING CHARGE Ordered / Back ordered : 1 / 0 EA			
0060	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Township Police Department 2400 Bethel Avenue Pennsauken, NJ 08109		2. Contact: Sgt. Chris Sulzbach	4. Date: 03/31/2015
		3. Phone Number: 856-488-0080	5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0051	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Alcotest instrument returned from outside evaluation and placed back in service.

10. Comments:
Return and Repair Form from Draeger indicates that the fuel cell and spacer plates were replaced.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Michael P. Gibson #6353
Name & Badge Number (Print)

Michael P. Gibson #6353 3/31/2015
Signature Date