

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03078	Calib. Date: 02/09/2022
Certification File No.:	03079	Calib. No.: 00044
Linearity File No.:	03080	Cert. Date: 02/09/2022
Solution File No.:	03084	Cert. No.: 00040
Sequential File No.:	03084	Lin. Date: 02/09/2022
		Lin. No.: 00039
		Soln. Date: 04/09/2022
		Soln. No.: 00361
		File Date: 04/09/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0656

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:38D		
Control 1 EC	0.100%	14:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39D		
Control 2 EC	0.098%	14:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:40D		
Control 3 EC	0.098%	14:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 04/09/2022