

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	01496	Calib. Date:	10/16/2012
Certification File No.:	01497	Cert. No.:	00020
Linearity File No.:	01498	Cert. Date:	10/16/2012
Solution File No.:	01530	Lin. No.:	00017
Sequential File No.:	01530	Lin. Date:	10/16/2012
		Soln. No.:	00142
		Soln. Date:	12/06/2012
		File Date:	12/06/2012

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0341
Control Solution %:	0.100%			Expires:	03/02/2014
Solution Control Lot:	12C100			Bottle No.:	1186

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:14S		
Control 1 EC	0.098%	03:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	03:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15S		
Control 2 EC	0.097%	03:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16S		
Control 3 EC	0.097%	03:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____



149

Badge No.: 149

Date: 12/06/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01496	Calib. Date: 10/16/2012
Certification File No.:	01497	Calib. No.: 00020
Linearity File No.:	01498	Cert. Date: 10/16/2012
Solution File No.:	01512	Cert. No.: 00017
Sequential File No.:	01512	Lin. Date: 10/16/2012
		Lin. No.: 00017
		Soln. Date: 11/20/2012
		Soln. No.: 00141
		File Date: 11/20/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1198

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:55S		
Control 1 EC	0.100%	16:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:57S		
Control 2 EC	0.100%	16:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:58S		
Control 3 EC	0.100%	16:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Badge No.: 190

Date: 11/20/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01487	Cert. No.: 00016
Sequential File No.:	01487	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 09/28/2012
		Soln. No.: 00139
		File Date: 09/28/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1181

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:51D		
Control 1 EC	0.097%	19:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	19:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:53D		
Control 2 EC	0.096%	19:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	19:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54D		
Control 3 EC	0.097%	19:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	19:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Handwritten signature: [Signature] #190

Badge No.: 190

Date: 09/28/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01480	Cert. No.: 00016
Sequential File No.:	01480	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 09/13/2012
		Soln. No.: 00138
		File Date: 09/13/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1199

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:23D		
Control 1 EC	0.097%	06:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.095%	06:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25D		
Control 2 EC	0.097%	06:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:26D		
Control 3 EC	0.097%	06:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

FABI P. 187

Badge No.: 187

Date: 09/13/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01478	Cert. No.: 00016
Sequential File No.:	01478	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 09/09/2012
		Soln. No.: 00137
		File Date: 09/09/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1195

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:19D		
Control 1 EC	0.099%	05:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:20D		
Control 2 EC	0.098%	05:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:21D		
Control 3 EC	0.098%	05:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	05:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:22D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature:  149

Badge No.: 149

Date: 09/09/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01472	Cert. No.: 00016
Sequential File No.:	01472	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 09/03/2012
		Soln. No.: 00136
		File Date: 09/03/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1187

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:00D		
Control 1 EC	0.099%	19:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:01D		
Control 2 EC	0.098%	19:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	19:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:02D		
Control 3 EC	0.098%	19:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

[Handwritten Signature: G. Henkel]

[Handwritten: 149]

Badge No.: 149

Date: 09/03/2012

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 01355 Calib. Date: 05/01/2012 Calib. No.: 00019
Certification File No.: 01356 Cert. Date: 05/01/2012 Cert. No.: 00016
Linearity File No.: 01357 Lin. Date: 05/01/2012 Lin. No.: 00016
Solution File No.: 01464 Soln. Date: 08/20/2012 Soln. No.: 00135
Sequential File No.: 01464 File Date: 08/20/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 03/02/2014
Solution Control Lot: 12C100 Bottle No.: 1177

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:21D		
Control 1 EC	0.097%	06:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	06:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:22D		
Control 2 EC	0.096%	06:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		
Control 3 EC	0.096%	06:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	06:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:25D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: 

Badge No.: 190

Date: 08/20/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP, P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01454	Cert. No.: 00016
Sequential File No.:	01454	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 08/11/2012
		Soln. No.: 00134
		File Date: 08/11/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1188

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:23D		
Control 1 EC	0.099%	22:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:24D		
Control 2 EC	0.097%	22:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:25D		
Control 3 EC	0.098%	22:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 08/11/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01445	Cert. No.: 00016
Sequential File No.:	01445	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 08/03/2012
		Soln. No.: 00133
		File Date: 08/03/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1190

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:13D		
Control 1 EC	0.100%	05:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:15D		
Control 2 EC	0.099%	05:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:16D		
Control 3 EC	0.099%	05:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: *Ple M Henkel 149*

Badge No.: 149

Date: 08/03/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01413	Lin. No.: 00016
Sequential File No.:	01413	Lin. Date: 05/01/2012
		Soln. No.: 00131
		Soln. Date: 07/02/2012
		File Date: 07/02/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1179

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:41D		
Control 1 EC	0.100%	13:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		
Control 2 EC	0.099%	13:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:43D		
Control 3 EC	0.099%	13:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Am J #190

Badge No.: 190

Date: 07/02/2012

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 01355 Calib. Date: 05/01/2012 Calib. No.: 00019
Certification File No.: 01356 Cert. Date: 05/01/2012 Cert. No.: 00016
Linearity File No.: 01357 Lin. Date: 05/01/2012 Lin. No.: 00016
Solution File No.: 01399 Soln. Date: 06/16/2012 Soln. No.: 00130
Sequential File No.: 01399 File Date: 06/16/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 03/02/2014
Solution Control Lot: 12C100 Bottle No.: 1184

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:02D		
Control 1 EC	0.098%	02:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:04D		
Control 2 EC	0.098%	02:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	02:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:05D		
Control 3 EC	0.097%	02:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	02:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:06D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature:  149

Badge No.: 149

Date: 06/16/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01388	Cert. No.: 00016
Sequential File No.:	01388	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 06/03/2012
		Soln. No.: 00129
		File Date: 06/03/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 1185

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:07D		
Control 1 EC	0.098%	02:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:08D		
Control 2 EC	0.099%	02:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:09D		
Control 3 EC	0.099%	02:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

Handwritten signature: G. Henkel 149

Badge No.: 149

Date: 06/03/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01355	Calib. Date: 05/01/2012
Certification File No.:	01356	Calib. No.: 00019
Linearity File No.:	01357	Cert. Date: 05/01/2012
Solution File No.:	01374	Cert. No.: 00016
Sequential File No.:	01374	Lin. Date: 05/01/2012
		Lin. No.: 00016
		Soln. Date: 05/13/2012
		Soln. No.: 00128
		File Date: 05/13/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0176

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:29D		
Control 1 EC	0.099%	22:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30D		
Control 2 EC	0.098%	22:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:31D		
Control 3 EC	0.099%	22:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr Jm #190

Badge No.: 190

Date: 05/13/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01249	Calib. Date: 01/26/2012
Certification File No.:	01250	Calib. No.: 00018
Linearity File No.:	01251	Cert. Date: 01/26/2012
Solution File No.:	01351	Cert. No.: 00015
Sequential File No.:	01351	Lin. Date: 01/26/2012
		Lin. No.: 00015
		Soln. Date: 04/27/2012
		Soln. No.: 00126
		File Date: 04/27/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0197

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:11D		
Control 1 EC	0.101%	01:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13D		
Control 2 EC	0.101%	01:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	01:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:14D		
Control 3 EC	0.101%	01:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

M J #190

Badge No.: 190

Date: 04/27/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01249	Calib. Date: 01/26/2012
Certification File No.:	01250	Calib. No.: 00018
Linearity File No.:	01251	Cert. Date: 01/26/2012
Solution File No.:	01331	Cert. No.: 00015
Sequential File No.:	01331	Lin. Date: 01/26/2012
		Lin. No.: 00015
		Soln. Date: 04/05/2012
		Soln. No.: 00125
		File Date: 04/05/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0200

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:29D		
Control 1 EC	0.102%	20:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:31D		
Control 2 EC	0.101%	20:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	20:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:32D		
Control 3 EC	0.101%	20:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Badge No.: 190

Date: 04/05/2012

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 01249 Calib. Date: 01/26/2012 Calib. No.: 00018
Certification File No.: 01250 Cert. Date: 01/26/2012 Cert. No.: 00015
Linearity File No.: 01251 Lin. Date: 01/26/2012 Lin. No.: 00015
Solution File No.: 01312 Soln. Date: 03/21/2012 Soln. No.: 00124
Sequential File No.: 01312 File Date: 03/21/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0200

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:57S		
Control 1 EC	0.100%	15:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58S		
Control 2 EC	0.100%	15:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00S		
Control 3 EC	0.101%	16:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:01S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: 

Badge No.: 190

Date: 03/21/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01249	Calib. Date: 01/26/2012
Certification File No.:	01250	Calib. No.: 00018
Linearity File No.:	01251	Cert. Date: 01/26/2012
Solution File No.:	01292	Lin. Date: 01/26/2012
Sequential File No.:	01292	Lin. No.: 00015
		Soln. Date: 02/28/2012
		Soln. No.: 00123
		File Date: 02/28/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0187

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:26S		
Control 1 EC	0.102%	07:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28S		
Control 2 EC	0.102%	07:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	07:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:29S		
Control 3 EC	0.102%	07:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMPSON

First Name: SHAUN

MI: .

Signature: _____



Badge No.: 142

Date: 02/28/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01249	Calib. Date: 01/26/2012
Certification File No.:	01250	Calib. No.: 00018
Linearity File No.:	01251	Cert. Date: 01/26/2012
Solution File No.:	01270	Cert. No.: 00015
Sequential File No.:	01270	Lin. Date: 01/26/2012
		Lin. No.: 00015
		Soln. Date: 02/17/2012
		Soln. No.: 00122
		File Date: 02/17/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0193

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:36S		
Control 1 EC	0.103%	02:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:37S		
Control 2 EC	0.101%	02:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:39S		
Control 3 EC	0.101%	02:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

[Handwritten Signature: G. Henkel 149]

Badge No.: 149

Date: 02/17/2012

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 01119 Calib. Date: 07/28/2011 Calib. No.: 00017
Certification File No.: 01120 Cert. Date: 07/28/2011 Cert. No.: 00014
Linearity File No.: 01121 Lin. Date: 07/28/2011 Lin. No.: 00014
Solution File No.: 01236 Soln. Date: 01/07/2012 Soln. No.: 00120
Sequential File No.: 01236 File Date: 01/07/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0182

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:04S		
Control 1 EC	0.102%	00:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.104%	00:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06S		
Control 2 EC	0.102%	00:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	00:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:07S		
Control 3 EC	0.102%	00:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MOLES

First Name: VITO

MI: M

Signature: _____



Badge No.: 171

Date: 01/07/2012