

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	03078	Calib. Date:	02/09/2022
Certification File No.:	03079	Cert. No.:	00044
Linearity File No.:	03080	Cert. Date:	02/09/2022
Solution File No.:	03086	Lin. Date:	02/09/2022
Sequential File No.:	03086	Soln. Date:	05/14/2022
		File Date:	05/14/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0341
Solution Control Lot:	21270	Expires:	07/21/2023
		Bottle No.:	0155

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:43D		
Control 1 EC	0.100%	21:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:45D		
Control 2 EC	0.098%	21:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:46D		
Control 3 EC	0.098%	21:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ADAIR

First Name: CRAIG

MI: M

Signature: _____

Sgt. C LR # 200

Badge No.: 200

Date: 05/14/2022

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: PENNSAUKEN TWSP. P.D. Serial No.: ARUM-0051
Calibration File No.: 03078 Calib. Date: 02/09/2022 Calib. No.: 00044
Certification File No.: 03079 Cert. Date: 02/09/2022 Cert. No.: 00040
Linearity File No.: 03080 Lin. Date: 02/09/2022 Lin. No.: 00039
Solution File No.: 03089 Soln. Date: 06/11/2022 Soln. No.: 00363
Sequential File No.: 03089 File Date: 06/11/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0097

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:06D		
Control 1 EC	0.098%	20:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	20:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:08D		
Control 2 EC	0.097%	20:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	20:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09D		
Control 3 EC	0.097%	20:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ADAIR

First Name: CRAIG

MI: M

Signature: _____

Sgt-C ADAIR #200

Badge No.: 200

Date: 06/11/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	03090	Calib. Date:	07/11/2022
Certification File No.:	03091	Cert. No.:	00045
Linearity File No.:	03092	Cert. Date:	07/11/2022
Solution File No.:	03093	Lin. Date:	07/11/2022
Sequential File No.:	03093	Soln. Date:	07/11/2022
		File Date:	07/11/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	21270	Expires:	07/21/2023
		Bottle No.:	0156

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:21D		
Control 1 EC	0.100%	11:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		
Control 2 EC	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 3 EC	0.099%	11:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJP2-142 MRW

Changed By:

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature: Sgt.  #2028

Badge No.: 7078

Date: 07/11/2022