

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03067	Calib. Date: 09/20/2021
Certification File No.:	03068	Calib. No.: 00043
Linearity File No.:	03069	Cert. Date: 09/20/2021
Solution File No.:	03073	Cert. No.: 00039
Sequential File No.:	03073	Lin. Date: 09/20/2021
		Lin. No.: 00038
		Soln. Date: 12/18/2021
		Soln. No.: 00356
		File Date: 12/18/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0011

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:17S		
Control 1 EC	0.101%	04:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18S		
Control 2 EC	0.100%	04:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19S		
Control 3 EC	0.100%	04:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	04:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: TIMOTHY

MI: -

Signature: _____

Handwritten signature: Mireles 219

Badge No.: 219

Date: 12/18/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
 Location: PENNSAUKEN TWSP. P.D.
 Calibration File No.: 02851 Calib. Date: 09/20/2021 Calib. No.: 00041
 Certification File No.: 02852 Cert. Date: 09/20/2021 Cert. No.: 00037
 Linearity File No.: 02853 Lin. Date: 09/20/2021 Lin. No.: 00037
 Solution File No.: 02867 Soln. Date: 12/18/2021 Soln. No.: 00313
 Sequential File No.: 02867 File Date: 12/18/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
 Control Solution %: 0.100% Expires: 07/21/2023
 Solution Control Lot: 21270 Bottle No.: 0549

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:16S		
Control 1 EC	0.103%	04:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18S		
Control 2 EC	0.100%	04:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	04:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19S		
Control 3 EC	0.101%	04:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	04:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: TIMOTHY

MI: -

Signature: _____

T.M. Mireles #219

Badge No.: 219

Date: 12/18/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03067	Calib. Date: 09/20/2021
Certification File No.:	03068	Calib. No.: 00043
Linearity File No.:	03069	Cert. Date: 09/20/2021
Solution File No.:	03075	Cert. No.: 00039
Sequential File No.:	03075	Lin. Date: 09/20/2021
		Lin. No.: 00038
		Soln. Date: 12/21/2021
		Soln. No.: 00357
		File Date: 12/21/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0099

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:12S		
Control 1 EC	0.102%	13:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		
Control 2 EC	0.100%	13:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15S		
Control 3 EC	0.100%	13:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	13:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: PDuJP2-142 BMG

Changed By:

Last Name: GAMBONE

First Name: BRIAN

MI: M

Signature: Sgt. Brian Gambone #7029

Badge No.: 7029

Date: 12/21/2021