

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03037	Calib. Date: 12/17/2020
Certification File No.:	03038	Calib. No.: 00041
Linearity File No.:	03039	Cert. Date: 12/17/2020
Solution File No.:	03048	Cert. No.: 00037
Sequential File No.:	03048	Lin. Date: 12/17/2020
		Lin. No.: 00036
		Soln. Date: 03/07/2021
		Soln. No.: 00347
		File Date: 03/07/2021

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %:	0.100%		Expires: 09/02/2022
Solution Control Lot:	20390		Bottle No.: 0526

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:00S		
Control 1 EC	0.099%	09:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02S		
Control 2 EC	0.097%	09:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03S		
Control 3 EC	0.097%	09:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____

Badge No.: 187

Date: 03/07/2021

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.		
Calibration File No.: 03037	Calib. Date: 12/17/2020	Calib. No.: 00041
Certification File No.: 03038	Cert. Date: 12/17/2020	Cert. No.: 00037
Linearity File No.: 03039	Lin. Date: 12/17/2020	Lin. No.: 00036
Solution File No.: 03054	Soln. Date: 04/01/2021	Soln. No.: 00348
Sequential File No.: 03054	File Date: 04/01/2021	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUN S3-0340
Control Solution %: 0.100%		Expires: 09/02/2022
Solution Control Lot: 20390		Bottle No.: 0473

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:54S		
Control 1 EC	0.098%	21:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:56S		
Control 2 EC	0.096%	21:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:57S		
Control 3 EC	0.096%	21:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. of Mireles # 219

Badge No.: 219

Date: 04/01/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	03055	Calib. Date: 04/22/2021
Certification File No.:	03056	Calib. No.: 00042
Linearity File No.:	03057	Cert. Date: 04/22/2021
Solution File No.:	03061	Cert. No.: 00038
Sequential File No.:	03061	Lin. Date: 04/22/2021
		Lin. No.: 00037
		Soln. Date: 05/30/2021
		Soln. No.: 00350
		File Date: 05/30/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 0471

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:20D		
Control 1 EC	0.099%	08:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21D		
Control 2 EC	0.097%	08:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23D		
Control 3 EC	0.098%	08:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____

Badge No.: 187

Date: 05/30/2021

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.	
Calibration File No.: 03055	Calib. Date: 04/22/2021
Certification File No.: 03056	Cert. No.: 00042
Linearity File No.: 03057	Cert. Date: 04/22/2021
Solution File No.: 03062	Lin. No.: 00037
Sequential File No.: 03062	Soln. Date: 07/01/2021
	Soln. No.: 00351
	File Date: 07/01/2021

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUN S3-0338
Control Solution %: 0.100%		Expires: 03/17/2023
Solution Control Lot: 21090		Bottle No.: 0737

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:35D		
Control 1 EC	0.101%	09:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37D		
Control 2 EC	0.099%	09:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 3 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SEXTON

First Name: RYAN

MI: G

Signature: _____

RYAN SEXTON # 206

Badge No.: 206

Date: 07/01/2021

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.		
Calibration File No.: 03055	Calib. Date: 04/22/2021	Calib. No.: 00042
Certification File No.: 03056	Cert. Date: 04/22/2021	Cert. No.: 00038
Linearity File No.: 03057	Lin. Date: 04/22/2021	Lin. No.: 00037
Solution File No.: 03064	Soln. Date: 08/08/2021	Soln. No.: 00352
Sequential File No.: 03064	File Date: 08/08/2021	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUN S3-0338
Control Solution %: 0.100%		Expires: 09/09/2021
Solution Control Lot: 19240		Bottle No.: 1285

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:25D		
Control 1 EC	0.098%	08:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:26D		
Control 2 EC	0.097%	08:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:28D		
Control 3 EC	0.097%	08:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____



Badge No.: 187

Date: 08/08/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	03067	Calib. Date:	09/20/2021
Certification File No.:	03068	Cert. Date:	09/20/2021
Linearity File No.:	03069	Lin. Date:	09/20/2021
Solution File No.:	03072	Soln. Date:	11/14/2021
Sequential File No.:	03072	File Date:	11/14/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	0539

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:27S		
Control 1 EC	0.098%	12:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		
Control 2 EC	0.097%	12:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30S		
Control 3 EC	0.098%	12:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SEXTON

First Name: RYAN

MI: G

Signature: *PH*  # 206

Badge No.: 206

Date: 11/14/2021