

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARUM-0066 |
| Location: | PENNSAUKEN TWSP. P.D. | |
| Calibration File No.: | 02799 | Calib. Date: 12/17/2020 |
| Certification File No.: | 02800 | Cert. No.: 00035 |
| Linearity File No.: | 02801 | Lin. Date: 12/17/2020 |
| Solution File No.: | 02820 | Soln. Date: 03/25/2021 |
| Sequential File No.: | 02820 | File Date: 03/25/2021 |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUN S3-0339 |
| Solution Control Lot: | 20390 | Expires: 09/02/2022 |
| | | Bottle No.: 0554 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:15S | | |
| Control 1 EC | 0.099% | 07:16S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 07:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:17S | | |
| Control 2 EC | 0.099% | 07:17S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 07:17S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:18S | | |
| Control 3 EC | 0.098% | 07:19S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 07:19S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:20S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____

Badge No.: 187

Date: 03/25/2021

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARUM-0066 |
| Location: | PENNSAUKEN TWSP. P.D. | |
| Calibration File No.: | 02799 | Calib. Date: 12/17/2020 |
| Certification File No.: | 02800 | Calib. No.: 00039 |
| Linearity File No.: | 02801 | Cert. Date: 12/17/2020 |
| Solution File No.: | 02823 | Cert. No.: 00035 |
| Sequential File No.: | 02823 | Lin. Date: 12/17/2020 |
| | | Lin. No.: 00035 |
| | | Soln. Date: 04/01/2021 |
| | | Soln. No.: 00304 |
| | | File Date: 04/01/2021 |

| | | | |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 | Serial No.: DDUN S3-0339 |
| Control Solution %: | 0.100% | | Expires: 09/02/2022 |
| Solution Control Lot: | 20390 | | Bottle No.: 0472 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 21:46S | | |
| Control 1 EC | 0.098% | 21:47S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 21:47S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:48S | | |
| Control 2 EC | 0.097% | 21:48S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 21:48S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:49S | | |
| Control 3 EC | 0.097% | 21:50S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 21:50S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:50S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. Mireles #219

Badge No.: 219

Date: 04/01/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02831 Calib. Date: 04/22/2021 Calib. No.: 00040
Certification File No.: 02832 Cert. Date: 04/22/2021 Cert. No.: 00036
Linearity File No.: 02833 Lin. Date: 04/22/2021 Lin. No.: 00036
Solution File No.: 02836 Soln. Date: 05/30/2021 Soln. No.: 00306
Sequential File No.: 02836 File Date: 05/30/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 03/17/2023
Solution Control Lot: 21090 Bottle No.: 0807

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 08:15D | | |
| Control 1 EC | 0.100% | 08:16D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 08:16D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:17D | | |
| Control 2 EC | 0.099% | 08:17D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 08:17D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:18D | | |
| Control 3 EC | 0.099% | 08:19D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 08:19D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:19D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____



Badge No.: 187

Date: 05/30/2021

Calibrating Unit

New Standard Solution Report

| | | | | | |
|-------------------------|-----------------------|--------------|------------|-------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARUM-0066 | | |
| Location: | PENNSAUKEN TWSP. P.D. | | | | |
| Calibration File No.: | 02831 | Calib. Date: | 04/22/2021 | Calib. No.: | 00040 |
| Certification File No.: | 02832 | Cert. Date: | 04/22/2021 | Cert. No.: | 00036 |
| Linearity File No.: | 02833 | Lin. Date: | 04/22/2021 | Lin. No.: | 00036 |
| Solution File No.: | 02839 | Soln. Date: | 07/01/2021 | Soln. No.: | 00307 |
| Sequential File No.: | 02839 | File Date: | 07/01/2021 | | |
| | | | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUN S3-0341 |
| Control Solution %: | 0.100% | | | Expires: | 03/17/2023 |
| Solution Control Lot: | 21090 | | | Bottle No.: | 0763 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:27D | | |
| Control 1 EC | 0.100% | 08:27D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 08:27D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:28D | | |
| Control 2 EC | 0.099% | 08:29D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 08:29D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:29D | | |
| Control 3 EC | 0.098% | 08:30D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 08:30D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:31D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SEXTON

First Name: RYAN

MI: G

Signature: _____

RYAN SEXTON #206

Badge No.: 206

Date: 07/01/2021

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARUM-0066 |
| Location: | PENNSAUKEN TWSP. P.D. | |
| Calibration File No.: | 02831 | Calib. Date: 04/22/2021 |
| Certification File No.: | 02832 | Calib. No.: 00040 |
| Linearity File No.: | 02833 | Cert. Date: 04/22/2021 |
| Solution File No.: | 02845 | Cert. No.: 00036 |
| Sequential File No.: | 02845 | Lin. Date: 04/22/2021 |
| | | Lin. No.: 00036 |
| | | Soln. Date: 08/08/2021 |
| | | Soln. No.: 00308 |
| | | File Date: 08/08/2021 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUN S3-0341 |
| Solution Control Lot: | 20390 | Expires: 09/02/2022 |
| | | Bottle No.: 0182 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:24D | | |
| Control 1 EC | 0.098% | 08:25D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 08:25D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:25D | | |
| Control 2 EC | 0.097% | 08:26D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 08:26D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:27D | | |
| Control 3 EC | 0.098% | 08:27D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 08:27D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:28D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____



Badge No.: 187

Date: 08/08/2021

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARUM-0066 |
| Location: | PENNSAUKEN TWSP. P.D. | |
| Calibration File No.: | 02851 | Calib. Date: 09/20/2021 |
| Certification File No.: | 02852 | Calib. No.: 00041 |
| Linearity File No.: | 02853 | Cert. Date: 09/20/2021 |
| Solution File No.: | 02858 | Cert. No.: 00037 |
| Sequential File No.: | 02858 | Lin. Date: 09/20/2021 |
| | | Lin. No.: 00037 |
| | | Soln. Date: 11/14/2021 |
| | | Soln. No.: 00311 |
| | | File Date: 11/14/2021 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUN S3-0339 |
| Solution Control Lot: | 20390 | Expires: 09/02/2022 |
| | | Bottle No.: 0238 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 12:34S | | |
| Control 1 EC | 0.000% | 12:35S | 33.9°C | CONTROL TEST FAILED |
| Control 1 IR | 0.000% | 12:35S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 12:35S | | |
| Control 2 EC | -.---% | 12:36S | 34.0°C | CTRL GAS SUPPLY |
| Control 2 IR | -.---% | 12:36S | 34.0°C | CTRL GAS SUPPLY |
| Ambient Air Blank | -.---% | --:-- | | |
| Control 3 EC | -.---% | --:-- | N/A | |
| Control 3 IR | -.---% | --:-- | N/A | |
| Ambient Air Blank | -.---% | --:-- | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SEXTON

First Name: RYAN

MI: G

Signature: _____

PH  #206

Badge No.: 206

Date: 11/14/2021

