

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Police Department 2400 Bethel Ave #1 Pennsauken Twp, NJ 08109	2. Contact: Sgt. Ray Kern 3. Phone Number: 856-488-0080	4. Date: 06/06/2023 5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUM-0051</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:

I arrived to Pennsauken PD to perform the semi-annual recalibration for their alcotest instrument. Upon my arrival, I observed alcotest ARUM-0051 cycle on and then off repeatedly. I attempted to power the instrument on and off using the power switch, but that had no effect and the instrument continued to power on and then off on its own.

10. Comments:

Operational condition of the instrument was assessed by viewing the instrument and observing it purge on and off.

The instrument was placed out of service and sent back to Draeger for evaluation.

Last Known Sequential File #: 3150

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Joshua B. Hall #7750
Name & Badge Number (Print)


Signature

06/06/2023
Date