


STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Police Department 2400 Bethel Ave Pennsauken Twp, NJ 08109	2. Contact: Ray Kern 3. Phone Number: 856-488-0080	4. Date: 12/04/2023 5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0051	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:
See Draeger Return and Repair form.


NEW JERSEY STATE POLICE
1921

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Joshua B. Hall #7750
Name & Badge Number (Print)


Signature

12/04/2023
Date



Delivery receipt



Customer no.
92560923

Number of report
308332889
Please reference on inquiries

Date of receiving Delivery receipt date
07/10/2023 09/28/2023

Consignee, 150049890
PENNSAUKEN TOWNSHIP POLICE DEPT
Sgt. Kern #181 (Traffic Division)
2400 BETHEL AVE
PENNSAUKEN NJ 08109-2767

Customer, 92560923
TOWNSHIP OF PENNSAUKEN
Walt Nicgorski
5605 N CRESCENT BLVD
PENNSAUKEN NJ 08110-1834

Your order

Date of order:
06/28/2023
Purch.ord.no.:
23-03558

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
000	ALCOTEST 7110 MK III USA	8314246	1148377398	ARUM-0051
	REPAIR ARUM-0051			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road, Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no. Number of report Date of receiving
 92560923 308332889 07/10/2023
 Please reference on inquiries

Customer
 TOWNSHIP OF PENNSAUKEN
 5605 N CRESCENT BLVD
 PENNSAUKEN NJ 08110-1834

Consignee
 PENNSAUKEN TOWNSHIP POLICE DEPT
 2400 BETHEL AVE
 PENNSAUKEN NJ 08109-2767

Your order

Date of order:
 06/28/2023
 Your reference:
 23-03558

NL-Text

Your contact person

Service Support-IDT
 800-437-2437 OPT 5
 SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 148969336 REPAIR ARUM-0051 Eq.No. : 1148377398 SN: ARUM-0051 Material: 8314246 ALCOTEST 7110 MK III USA Inv.No. : Location: REPAIR ARUM-0051 REPAIR ARUM-0051 R004 Repair time - Workshop REPAIR ARUM-0051 Testing and QC completed after power supply replacement.
	1.0 H	

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Township Of Pennsauken , Pennsauken , NJ

Date Received: 07/10/2023

Date given to service: 07/10/2023

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0051

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	.5	
4412036	Power Supply	1	

Repair Notes: Replaced defective power supply.

CAL W/QC AND OPS CHECK

Service Technician: GR / MB

Date: 09/21/2023

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Police Department 2400 Bethel Ave #1 Pennsauken Twp, NJ 08109	2. Contact: Sgt. Ray Kern	4. Date: 06/06/2023
	3. Phone Number: 856-488-0080	5. County: Camden
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUM-0051</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:

I arrived to Pennsauken PD to perform the semi-annual recalibration for their alcotest instrument. Upon my arrival, I observed alcotest ARUM-0051 cycle on and then off repeatedly. I attempted to power the instrument on and off using the power switch, but that had no effect and the instrument continued to power on and then off on its own.

10. Comments:

Operational condition of the instrument was assessed by viewing the instrument and observing it purge on and off.


The instrument was placed out of service and sent back to Draeger for evaluation.

Last Known Sequential File #: 3150

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Joshua B. Hall #7750
Name & Badge Number (Print)


Signature

 06/06/2023
Date