

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02905	Calib. Date: 08/26/2019
Certification File No.:	02906	Calib. No.: 00038
Linearity File No.:	02907	Cert. Date: 08/26/2019
Solution File No.:	02967	Cert. No.: 00034
Sequential File No.:	02967	Lin. Date: 08/26/2019
		Lin. No.: 00033
		Soln. Date: 12/25/2019
		Soln. No.: 00328
		File Date: 12/25/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	19150	Expires: 06/21/2021
		Bottle No.: 0252

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:52S		
Control 1 EC	0.097%	08:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54S		
Control 2 EC	0.095%	08:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56S		
Control 3 EC	0.096%	08:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

G M Henkel 149

Badge No.: 149

Date: 12/25/2019

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02905	Calib. Date: 08/26/2019
Certification File No.:	02906	Calib. No.: 00038
Linearity File No.:	02907	Cert. Date: 08/26/2019
Solution File No.:	02966	Cert. No.: 00034
Sequential File No.:	02966	Lin. Date: 08/26/2019
		Lin. No.: 00033
		Soln. Date: 12/24/2019
		Soln. No.: 00327
		File Date: 12/24/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0310

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:25S		
Control 1 EC	-.---%	07:25S	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	07:25S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PENNINGTON	First Name: MICHAEL	MI: J
Signature: 		Badge No.: 191
		Date: 12/24/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02905	Calib. Date:	08/26/2019
Certification File No.:	02906	Cert. No.:	00038
Linearity File No.:	02907	Cert. Date:	08/26/2019
Solution File No.:	02961	Lin. Date:	08/26/2019
Sequential File No.:	02961	Lin. No.:	00033
		Soln. Date:	12/15/2019
		Soln. No.:	00326
		File Date:	12/15/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	19150	Expires:	06/10/2021
		Bottle No.:	0251

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:23S		
Control 1 EC	0.100%	20:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:24S		
Control 2 EC	0.097%	20:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:25S		
Control 3 EC	0.097%	20:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. Mireles #219

Badge No.: 219

Date: 12/15/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02905	Calib. Date: 08/26/2019
Certification File No.:	02906	Calib. No.: 00038
Linearity File No.:	02907	Cert. Date: 08/26/2019
Solution File No.:	02954	Cert. No.: 00034
Sequential File No.:	02954	Lin. Date: 08/26/2019
		Lin. No.: 00033
		Soln. Date: 12/08/2019
		Soln. No.: 00325
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0199

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:53S		
Control 1 EC	0.099%	04:53S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:54S		
Control 2 EC	0.097%	04:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:56S		
Control 3 EC	0.097%	04:56S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

Handwritten signature: T. Mireles #219

Badge No.: 219

Date: 12/08/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
 Location: PENNSAUKEN TWSP. P.D.
 Calibration File No.: 02905 Calib. Date: 08/26/2019 Calib. No.: 00038
 Certification File No.: 02906 Cert. Date: 08/26/2019 Cert. No.: 00034
 Linearity File No.: 02907 Lin. Date: 08/26/2019 Lin. No.: 00033
 Solution File No.: 02946 Soln. Date: 11/25/2019 Soln. No.: 00324
 Sequential File No.: 02946 File Date: 11/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
 Control Solution %: 0.100% Expires: 06/10/2021
 Solution Control Lot: 19150 Bottle No.: 0323

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:29S		
Control 1 EC	0.100%	15:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31S		
Control 2 EC	0.098%	15:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32S		
Control 3 EC	0.099%	15:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: PETER

MI: E

Signature: _____

Badge No.: 187

Date: 11/25/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02905	Calib. Date: 08/26/2019
Certification File No.:	02906	Calib. No.: 00038
Linearity File No.:	02907	Cert. Date: 08/26/2019
Solution File No.:	02938	Cert. No.: 00034
Sequential File No.:	02938	Lin. Date: 08/26/2019
		Lin. No.: 00033
		Soln. Date: 11/16/2019
		Soln. No.: 00323
		File Date: 11/16/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0322

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:25S		
Control 1 EC	0.100%	01:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:26S		
Control 2 EC	0.099%	01:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:28S		
Control 3 EC	0.099%	01:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. Mireles #219

Badge No.: 219

Date: 11/16/2019

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02905 Calib. Date: 08/26/2019 Calib. No.: 00038
Certification File No.: 02906 Cert. Date: 08/26/2019 Cert. No.: 00034
Linearity File No.: 02907 Lin. Date: 08/26/2019 Lin. No.: 00033
Solution File No.: 02927 Soln. Date: 11/06/2019 Soln. No.: 00322
Sequential File No.: 02927 File Date: 11/06/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0309

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:02S		
Control 1 EC	0.100%	22:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:04S		
Control 2 EC	0.098%	22:04S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:05S		
Control 3 EC	0.098%	22:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: _____

M. Mireles #219

Badge No.: 219

Date: 11/06/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02905	Calib. Date: 08/26/2019
Certification File No.:	02906	Calib. No.: 00038
Linearity File No.:	02907	Cert. Date: 08/26/2019
Solution File No.:	02914	Cert. No.: 00034
Sequential File No.:	02914	Lin. Date: 08/26/2019
		Lin. No.: 00033
		Soln. Date: 10/26/2019
		Soln. No.: 00321
		File Date: 10/26/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0253

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:22D		
Control 1 EC	0.100%	00:23D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24D		
Control 2 EC	0.099%	00:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:25D		
Control 3 EC	0.099%	00:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: T

MI: -

Signature: PH. Mireles #219

Badge No.: 219

Date: 10/26/2019

Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02905 Calib. Date: 08/26/2019 Calib. No.: 00038
Certification File No.: 02906 Cert. Date: 08/26/2019 Cert. No.: 00034
Linearity File No.: 02907 Lin. Date: 08/26/2019 Lin. No.: 00033
Solution File No.: 02909 Soln. Date: 10/11/2019 Soln. No.: 00320
Sequential File No.: 02909 File Date: 10/11/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0136

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:15D		
Control 1 EC	0.099%	10:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		
Control 2 EC	0.098%	10:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 3 EC	0.098%	10:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL First Name: GERALD MI: F.
Signature: *Lt M Henkel 149* Badge No.: 149
Date: 10/11/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02904	Cert. No.: 00033
Sequential File No.:	02904	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 08/05/2019
		Soln. No.: 00318
		File Date: 08/05/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1159

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:58D		
Control 1 EC	0.093%	22:59D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.096%	22:59D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:00D		
Control 2 EC	0.093%	23:00D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.095%	23:00D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:01D		
Control 3 EC	0.093%	23:01D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.096%	23:01D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	23:02D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: PTL. Bristow #227

Badge No.: 227

Date: 08/05/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02903	Cert. No.: 00033
Sequential File No.:	02903	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 08/03/2019
		Soln. No.: 00317
		File Date: 08/03/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1192

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:10D		
Control 1 EC	0.094%	04:11D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.097%	04:11D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	04:12D		
Control 2 EC	0.093%	04:12D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.097%	04:12D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	04:13D		
Control 3 EC	0.094%	04:13D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.097%	04:13D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	04:14D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTC BRT #227

Badge No.: 227

Date: 08/03/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02901	Cert. No.: 00033
Sequential File No.:	02901	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 07/24/2019
		Soln. No.: 00316
		File Date: 07/24/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1287

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:05D		
Control 1 EC	0.096%	00:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		
Control 2 EC	0.096%	00:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	00:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:07D		
Control 3 EC	0.096%	00:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____

Badge No.: 191

Date: 07/24/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02859	Calib. Date:	04/11/2019
Certification File No.:	02860	Cert. Date:	04/11/2019
Linearity File No.:	02861	Lin. Date:	04/11/2019
Solution File No.:	02899	Soln. Date:	07/19/2019
Sequential File No.:	02899	File Date:	07/19/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	1054

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:28D		
Control 1 EC	0.101%	15:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29D		
Control 2 EC	0.100%	15:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30D		
Control 3 EC	0.099%	15:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBONE

First Name: BRIAN

MI: M

Signature: TPR. I B M #7029

Badge No.: 7029

Date: 07/19/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02859	Calib. Date:	04/11/2019
Certification File No.:	02860	Cert. Date:	04/11/2019
Linearity File No.:	02861	Lin. Date:	04/11/2019
Solution File No.:	02893	Soln. Date:	06/02/2019
Sequential File No.:	02893	File Date:	06/02/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	1157

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:39D		
Control 1 EC	0.098%	02:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D		
Control 2 EC	0.097%	02:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	02:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:41D		
Control 3 EC	0.097%	02:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTL. Bristow #227

Badge No.: 227

Date: 06/02/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02859 Calib. Date: 04/11/2019 Calib. No.: 00037
Certification File No.: 02860 Cert. Date: 04/11/2019 Cert. No.: 00033
Linearity File No.: 02861 Lin. Date: 04/11/2019 Lin. No.: 00032
Solution File No.: 02888 Soln. Date: 05/26/2019 Soln. No.: 00313
Sequential File No.: 02888 File Date: 05/26/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0931

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:36D		
Control 1 EC	0.097%	05:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:37D		
Control 2 EC	0.096%	05:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:39D		
Control 3 EC	0.096%	05:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HURLEY

First Name: R

MI: J

Signature: _____

PH. R. H. #205

Badge No.: 205

Date: 05/26/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Cert. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02882	Lin. No.: 00032
Sequential File No.:	02882	Soln. Date: 05/18/2019
		Soln. No.: 00312
		File Date: 05/18/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0959

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:02D		
Control 1 EC	0.098%	00:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:03D		
Control 2 EC	0.097%	00:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:04D		
Control 3 EC	0.097%	00:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 05/18/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02859 Calib. Date: 04/11/2019 Calib. No.: 00037
Certification File No.: 02860 Cert. Date: 04/11/2019 Cert. No.: 00033
Linearity File No.: 02861 Lin. Date: 04/11/2019 Lin. No.: 00032
Solution File No.: 02875 Soln. Date: 05/04/2019 Soln. No.: 00311
Sequential File No.: 02875 File Date: 05/04/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 1403

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:15D		
Control 1 EC	0.098%	03:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17D		
Control 2 EC	0.097%	03:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18D		
Control 3 EC	0.097%	03:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr Joe #216

Badge No.: 216

Date: 05/04/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02865	Cert. No.: 00033
Sequential File No.:	02865	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 04/13/2019
		Soln. No.: 00310
		File Date: 04/13/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:53D		
Control 1 EC	0.100%	01:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:54D		
Control 2 EC	0.097%	01:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:55D		
Control 3 EC	0.097%	01:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 191

Date: 04/13/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Cert. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02853	Lin. No.: 00031
Sequential File No.:	02853	Soln. Date: 03/30/2019
		Soln. No.: 00308
		File Date: 03/30/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0935

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:35S		
Control 1 EC	0.099%	05:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36S		
Control 2 EC	0.097%	05:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:38S		
Control 3 EC	0.097%	05:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: _____

PTL. Bristow #227

Badge No.: 227

Date: 03/30/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02804 Calib. Date: 10/24/2018 Calib. No.: 00036
Certification File No.: 02805 Cert. Date: 10/24/2018 Cert. No.: 00032
Linearity File No.: 02806 Lin. Date: 10/24/2018 Lin. No.: 00031
Solution File No.: 02846 Soln. Date: 03/20/2019 Soln. No.: 00307
Sequential File No.: 02846 File Date: 03/20/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 1408

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:25S		
Control 1 EC	0.099%	02:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:26S		
Control 2 EC	0.097%	02:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27S		
Control 3 EC	0.098%	02:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____



Badge No.: 216

Date: 03/20/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02841	Cert. No.: 00032
Sequential File No.:	02841	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 02/23/2019
		Soln. No.: 00306
		File Date: 02/23/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0308

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:24S		
Control 1 EC	0.100%	01:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:26S		
Control 2 EC	0.098%	01:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:27S		
Control 3 EC	0.098%	01:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 191

Date: 02/23/2019



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02804	Calib. Date:	10/24/2018
Certification File No.:	02805	Cert. No.:	00036
Linearity File No.:	02806	Cert. Date:	10/24/2018
Solution File No.:	02835	Lin. Date:	10/24/2018
Sequential File No.:	02835	Lin. No.:	00031
		Soln. Date:	01/28/2019
		Soln. No.:	00305
		File Date:	01/28/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	18090	Expires:	03/13/2020
		Bottle No.:	0493

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:13S		
Control 1 EC	0.100%	12:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15S		
Control 2 EC	0.098%	12:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16S		
Control 3 EC	0.098%	12:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: _____

Sgt M Henkel 149

Badge No.: 149

Date: 01/28/2019

