

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02524 Calib. Date: 08/22/2016 Calib. No.: 00031
Certification File No.: 02525 Cert. Date: 08/22/2016 Cert. No.: 00027
Linearity File No.: 02526 Lin. Date: 08/22/2016 Lin. No.: 00026
Solution File No.: 02578 Soln. Date: 01/07/2017 Soln. No.: 00264
Sequential File No.: 02578 File Date: 01/07/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0203

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:15S		
Control 1 EC	0.102%	00:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	00:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16S		
Control 2 EC	0.101%	00:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18S		
Control 3 EC	0.101%	00:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	00:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

Pel P. Fabiani 01/07/2017

Badge No.: 187

Date: 01/07/2017

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02579 Calib. Date: 01/11/2017 Calib. No.: 00032
Certification File No.: 02580 Cert. Date: 01/11/2017 Cert. No.: 00028
Linearity File No.: 02581 Lin. Date: 01/11/2017 Lin. No.: 00027
Solution File No.: 02591 Soln. Date: 02/06/2017 Soln. No.: 00266
Sequential File No.: 02591 File Date: 02/06/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0222

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.100%	07:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54S		
Control 2 EC	0.098%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 3 EC	0.099%	07:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 02/06/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02602	Cert. No.: 00028
Sequential File No.:	02602	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 03/06/2017
		Soln. No.: 00267
		File Date: 03/06/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:07S		
Control 1 EC	0.099%	06:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:08S		
Control 2 EC	0.098%	06:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10S		
Control 3 EC	0.098%	06:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

[Handwritten Signature] #187

Badge No.: 187

Date: 03/06/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017 Calib. No.: 00032
Certification File No.:	02580	Cert. Date: 01/11/2017 Cert. No.: 00028
Linearity File No.:	02581	Lin. Date: 01/11/2017 Lin. No.: 00027
Solution File No.:	02608	Soln. Date: 04/08/2017 Soln. No.: 00268
Sequential File No.:	02608	File Date: 04/08/2017
Calibrating Unit:	WET	Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %:	0.100%	Expires: 10/28/2018
Solution Control Lot:	16330	Bottle No.: 0331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 2 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 3 EC	0.098%	09:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMPSON

First Name: OFFICER

MI: .

Signature: _____

R. Sampson # 142

Badge No.: 142

Date: 04/08/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02617	Cert. No.: 00028
Sequential File No.:	02617	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/07/2017
		Soln. No.: 00269
		File Date: 05/07/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0082

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:29D		
Control 1 EC	0.101%	00:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30D		
Control 2 EC	0.099%	00:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32D		
Control 3 EC	0.099%	00:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature:  187

Badge No.: 187

Date: 05/07/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02627	Cert. No.: 00028
Sequential File No.:	02627	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/24/2017
		Soln. No.: 00270
		File Date: 05/24/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0127

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:18D		
Control 1 EC	0.098%	03:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:20D		
Control 2 EC	0.097%	03:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:21D		
Control 3 EC	0.097%	03:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 05/24/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02579	Calib. Date:	01/11/2017
Certification File No.:	02580	Cert. No.:	00032
Linearity File No.:	02581	Calib. No.:	00028
Solution File No.:	02630	Lin. Date:	01/11/2017
Sequential File No.:	02630	Lin. No.:	00027
		Soln. Date:	05/27/2017
		Soln. No.:	00271
		File Date:	05/27/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	0457

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:54D		
Control 1 EC	0.097%	05:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:55D		
Control 2 EC	0.096%	05:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56D		
Control 3 EC	0.096%	05:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

Badge No.: 187

Date: 05/27/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02634	Cert. No.: 00028
Sequential File No.:	02634	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/29/2017
		Soln. No.: 00272
		File Date: 05/29/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:08D		
Control 1 EC	0.000%	03:09D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.000%	03:09D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	03:09D		
Control 2 EC	-.---%	03:10D	33.9°C	CTRL GAS SUPPLY
Control 2 IR	-.---%	03:10D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 05/29/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02636	Calib. Date: 07/11/2017
Certification File No.:	02637	Calib. No.: 00033
Linearity File No.:	02638	Cert. Date: 07/11/2017
Solution File No.:	02644	Cert. No.: 00029
Sequential File No.:	02644	Lin. Date: 07/11/2017
		Lin. No.: 00028
		Soln. Date: 08/04/2017
		Soln. No.: 00277
		File Date: 08/04/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0339
Control Solution %:	0.100%		Expires: 03/20/2019
Solution Control Lot:	17110		Bottle No.: 0898

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:34D		
Control 1 EC	0.097%	21:35D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:36D		
Control 2 EC	0.097%	21:36D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:37D		
Control 3 EC	0.098%	21:37D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 08/04/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02636	Calib. Date: 07/11/2017
Certification File No.:	02637	Calib. No.: 00033
Linearity File No.:	02638	Cert. Date: 07/11/2017
Solution File No.:	02649	Cert. No.: 00029
Sequential File No.:	02649	Lin. Date: 07/11/2017
		Lin. No.: 00028
		Soln. Date: 09/01/2017
		Soln. No.: 00278
		File Date: 09/01/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0880

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:53D		
Control 1 EC	0.101%	22:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:55D		
Control 2 EC	0.100%	22:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:56D		
Control 3 EC	0.100%	22:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature:  187

Badge No.: 187

Date: 09/01/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02636	Calib. Date: 07/11/2017
Certification File No.:	02637	Calib. No.: 00033
Linearity File No.:	02638	Cert. Date: 07/11/2017
Solution File No.:	02665	Cert. No.: 00029
Sequential File No.:	02665	Lin. Date: 07/11/2017
		Lin. No.: 00028
		Soln. Date: 09/30/2017
		Soln. No.: 00279
		File Date: 09/30/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 1112

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:12D		
Control 1 EC	0.101%	03:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	03:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:13D		
Control 2 EC	0.100%	03:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	03:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15D		
Control 3 EC	0.100%	03:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	03:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

[Handwritten Signature] #187

Badge No.: 187

Date: 09/30/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02636	Calib. Date: 07/11/2017
Certification File No.:	02637	Calib. No.: 00033
Linearity File No.:	02638	Cert. Date: 07/11/2017
Solution File No.:	02678	Cert. No.: 00029
Sequential File No.:	02678	Lin. Date: 07/11/2017
		Lin. No.: 00028
		Soln. Date: 10/23/2017
		Soln. No.: 00280
		File Date: 10/23/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUN S3-0339
Control Solution %:	0.100%		Expires: 03/20/2019
Solution Control Lot:	17110		Bottle No.: 0908

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:27D		
Control 1 EC	0.102%	14:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	14:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28D		
Control 2 EC	0.101%	14:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	14:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:29D		
Control 3 EC	0.101%	14:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 10/23/2017

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02636	Calib. Date: 07/11/2017
Certification File No.:	02637	Calib. No.: 00033
Linearity File No.:	02638	Cert. Date: 07/11/2017
Solution File No.:	02683	Cert. No.: 00029
Sequential File No.:	02683	Lin. Date: 07/11/2017
		Lin. No.: 00028
		Soln. Date: 11/17/2017
		Soln. No.: 00281
		File Date: 11/17/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0771

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:03S		
Control 1 EC	0.103%	21:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04S		
Control 2 EC	0.101%	21:05S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.104%	21:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05S		
Control 3 EC	0.101%	21:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	21:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Pr Joseph #216

Badge No.: 216

Date: 11/17/2017