

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Cert. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02832	Lin. No.: 00031
Sequential File No.:	02832	Soln. Date: 12/30/2018
		Soln. No.: 00304
		File Date: 12/30/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0085

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:30S		
Control 1 EC	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31S		
Control 2 EC	0.099%	08:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 3 EC	0.098%	08:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MIRELES

First Name: TIMOTHY

MI: I

Signature: _____



Badge No.: 219

Date: 12/30/2018