

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Pennsauken Police Department 2400 Bethel Ave. Pennsauken, NJ 08109</b>		2. Contact: <b>Sgt. Chris Sulzbach</b>	4. Date: <b>5/17/2016</b>
		3. Phone Number: <b>(856)488-0080 x2301</b>	5. County: <b>Camden</b>
6. Alcotest Instrument Serial Number: <b>ARUM-0066</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Complaint received of the instrument experiencing control test failures.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed .000% Electrochemical Sensor (EC) results on control tests.**  
  
**Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

STATE OF NEW JERSEY

Last Known Sequential File #: 02114

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Stephen M Grusemeyer #6909  
Name & Badge Number (Print)

*Stephen M Grusemeyer*  
Signature

Date  
**5/17/2016**

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Pennsauken Police Department 2400 Bethel Ave. Pennsauken, NJ 08109</b>		2. Contact: <b>Sgt. Chris Sulzbach</b>	4. Date: <b>7/19/2016</b>
		3. Phone Number: <b>(856)488-0080 x2301</b>	5. County: <b>Camden</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARUM-0066</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	

9. Reason for Service:  
**The alcotest instrument was returned from outside evaluation and placed back in service.**


10. Comments:  
**See Draeger Return and Repair form.**

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Matthew R. Watson #7078  
 Name & Badge Number (Print)

 #7078  
 Signature Date **7/19/2016**

# Packing Slip

Customer no    Order No    Order date  
 150056355    10851776    07/07/2016

Packing slip no.    Ship date  
 81029029    07/07/2016

Please reference on inquiries

Ship to  
 150049890  
 PENNSAUKEN TOWNSHIP  
 MUNICIPAL COMPLEX  
 ATTN: SGT. CHRIS SULZBACH  
 2400 BETHEL AVENUE  
 PENNSAUKEN, NJ 08109-2767  
 USA

Payer

Your Purchase Order Number  
 07/07/2016  
 SGT. CHRIS SULZBACH  
 Any questions? Please contact:  
 ROSEMARIE HUGHES

Customer  
 PENNSAUKEN POLICE DEPARTMENT  
 ATTN: SGT. CHRIS SULZBACH  
 CAMDEN COUNTY  
 2400 BETHEL AVENUE  
 PENNSAUKEN, NJ 08109-2792  
 USA

Your sales person  
 CHRISTOPHER FAIRCHILD  
 Phone: 866-385-5900  
 Fax: 972-929-1260

Delivery terms  
 FA  
 FREIGHT ALLOWED

Page 1 / 2 Complete delivery

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110:ARUM-0066 REC'D 5/24/16
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 1 / 0 EA
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.5 / 0.0 EA

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 536410  
 Pittsburgh, PA 15253-5906

# Packing Slip

Customer no    Order No    Order date  
150056355    10851776    07/07/2016

**Packing slip no.**    **Ship date**  
81029029    07/07/2016

Please reference on inquiries

Page    2 / 2    Complete delivery

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 1 / 0 EA			
0040	2 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 2 / 0 EA			
0050	1 EA	MPINS INSURANCE AND HANDLING CHARGE Ordered / Back ordered : 1 / 0 EA			
0060	1 EA	FREIGHT FREIGHT TO DRAEGER Ordered / Back ordered : 1 / 0 EA			
0070	1 EA	FREIGHT FREIGHT TO AGENCY Ordered / Back ordered : 1 / 0 EA			

# RETURN AND REPAIR FORM



## Customer Information

B: \_\_\_\_\_ S: \_\_\_\_\_

Company Name: Division of State Police, Pennsauken, NJ

Date Received: 05/24/2016 Date given to service: 05/24/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810 Serial Number: AR UM-0066  
 7110  8610 Printer Serial#: AR  
 7510  DT5000 Sim Serial#: DD  
 7410 Upper-half  9510 Probe Serial#: DD  
 7410 Whole

Warranty Expires: April 2013 (56355)

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJD-1781  
 CAL W/QC AND OPS CHECK

Service Technician: BC Date: 07/06/2016