

# Packing Slip

Customer no    Order No    Order date  
 150056355    10597888    01/10/2013

Packing slip no.    Ship date  
 80743195    01/10/2013

Please reference on inquiries

ship to

150049890  
 PENNSAUKEN TOWNSHIP  
 MUNICIPAL COMPLEX  
 2400 BETHEL AVENUE  
 PENNSAUKEN, NJ 08109  
 USA

bill to

Your Purchase Order Number

01/10/2013  
 RECERT/RECAL  
 Any questions? Please contact:  
 MARA HERRERA

Customer

PENNSAUKEN POLICE DEPARTMENT  
 ATTN: CHRIS SULZBACH  
 CAMDEN COUNTY  
 2400 BETHEL AVENUE  
 PENNSAUKEN, NJ 08109-2792  
 USA

Your sales person

Don Pouliot  
 Phone: 973-398-3228

Delivery terms

FA  
 FREIGHT ALLOWED

Page 1 / 2 Complete delivery

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		SIM: DDUNS3-0340 PROBE: DDUJP2-142			
0010	1 EA	MPCERT CERTIFICATION CHARGE SIMULATOR Ordered / Back ordered : 1 / 0 EA			
0020	1 EA	MPCALTP CAL CHARGE, SIM PROBE Ordered / Back ordered : 1 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150056355    10597888    01/10/2013

**Packing slip no.**    **Ship date**  
80743195    01/10/2013

Please reference on inquiries

Page    2 / 2    Complete delivery

Item#	sh. Quant.	Part no. Description			
0030	1 EA	12043 HOSE PUMP-SIM W/SIM-CUVETTE HOSE 2" Ordered / Back ordered : 1 / 0 EA			
0040	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA			

# Dräger

## ALCOTEST® 7110 TEMPERATURE PROBE

### CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDUJP2-142

Certification date:

01/09/13

Next Certification due:


01/09/14

Probe Value

104

Draeger Safety Diagnostics, Inc.  
Technical Service Department

SY



# Dräger

### CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34  
 Model: MARK IIA  
 Other: \_\_\_\_\_

Serial Number:

DDUN S3-0340

Certification Date

01/09/13

Technician

SY

Re-Certification Due Date

01/09/14

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 01911 Calib. Date: 10/16/2012 Calib. No.: 00020  
Certification File No.: 01912 Cert. Date: 10/16/2012 Cert. No.: 00017  
Linearity File No.: 01913 Lin. Date: 10/16/2012 Lin. No.: 00016  
Solution File No.: 01982 Soln. Date: 01/31/2013 Soln. No.: 00180  
Sequential File No.: 01982 File Date: 01/31/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 06/28/2014  
Solution Control Lot: 12F102 Bottle No.: 0851

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:18S		
Control 1 EC	0.103%	10:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		
Control 2 EC	0.100%	10:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21S		
Control 3 EC	0.100%	10:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUIPZ-142 (MFA)

**Changed By:**

Last Name: GIBSON

First Name: MICHAEL

MI: P

Signature: *Tr. Michael P. Gibson #6353*

Badge No.: 6353

Date: 01/31/2013