

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02524	Calib. Date:	08/22/2016
Certification File No.:	02525	Cert. No.:	00031
Linearity File No.:	02526	Cert. Date:	08/22/2016
Solution File No.:	02578	Lin. No.:	00027
Sequential File No.:	02578	Lin. Date:	08/22/2016
		Soln. No.:	00026
		Soln. Date:	01/07/2017
		File Date:	01/07/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	16330	Expires:	10/28/2018
		Bottle No.:	0203

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:15S		
Control 1 EC	0.102%	00:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	00:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16S		
Control 2 EC	0.101%	00:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18S		
Control 3 EC	0.101%	00:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	00:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

Pel P. Fabiani *187

Badge No.: 187

Date: 01/07/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02579 Calib. Date: 01/11/2017 Calib. No.: 00032
Certification File No.: 02580 Cert. Date: 01/11/2017 Cert. No.: 00028
Linearity File No.: 02581 Lin. Date: 01/11/2017 Lin. No.: 00027
Solution File No.: 02591 Soln. Date: 02/06/2017 Soln. No.: 00266
Sequential File No.: 02591 File Date: 02/06/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0222

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.100%	07:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54S		
Control 2 EC	0.098%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 3 EC	0.099%	07:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 02/06/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02602	Cert. No.: 00028
Sequential File No.:	02602	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 03/06/2017
		Soln. No.: 00267
		File Date: 03/06/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:07S		
Control 1 EC	0.099%	06:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:08S		
Control 2 EC	0.098%	06:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10S		
Control 3 EC	0.098%	06:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____

[Handwritten Signature] *187

Badge No.: 187

Date: 03/06/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02608	Cert. No.: 00028
Sequential File No.:	02608	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 04/08/2017
		Soln. No.: 00268
		File Date: 04/08/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0331

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 2 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 3 EC	0.098%	09:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMPSON

First Name: OFFICER

MI: .

Signature: _____

P. Sampson #142

Badge No.: 142

Date: 04/08/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02617	Cert. No.: 00028
Sequential File No.:	02617	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/07/2017
		Soln. No.: 00269
		File Date: 05/07/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0082

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:29D		
Control 1 EC	0.101%	00:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30D		
Control 2 EC	0.099%	00:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32D		
Control 3 EC	0.099%	00:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 05/07/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02627	Cert. No.: 00028
Sequential File No.:	02627	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/24/2017
		Soln. No.: 00270
		File Date: 05/24/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0127

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:18D		
Control 1 EC	0.098%	03:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:20D		
Control 2 EC	0.097%	03:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:21D		
Control 3 EC	0.097%	03:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 05/24/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02630	Cert. No.: 00028
Sequential File No.:	02630	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/27/2017
		Soln. No.: 00271
		File Date: 05/27/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0457

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:54D		
Control 1 EC	0.097%	05:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:55D		
Control 2 EC	0.096%	05:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56D		
Control 3 EC	0.096%	05:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



#187

Badge No.: 187

Date: 05/27/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02634	Cert. No.: 00028
Sequential File No.:	02634	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 05/29/2017
		Soln. No.: 00272
		File Date: 05/29/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:08D		
Control 1 EC	0.000%	03:09D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.000%	03:09D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	03:09D		
Control 2 EC	----%	03:10D	33.9°C	CTRL GAS SUPPLY
Control 2 IR	----%	03:10D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	---:--		
Control 3 EC	----%	---:--	N/A	
Control 3 IR	----%	---:--	N/A	
Ambient Air Blank	----%	---:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: _____



Badge No.: 187

Date: 05/29/2017