

Alcotest 7110 Calibration Record


Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02636 Calib. Date: 07/11/2017 Calib. No.: 00033
Certification File No.: 02580 Cert. Date: 01/11/2017 Cert. No.: 00028
Linearity File No.: 02581 Lin. Date: 01/11/2017 Lin. No.: 00027
Solution File No.: 02635 Soln. Date: 06/01/2017 Soln. No.: 00273
Sequential File No.: 02636 File Date: 07/11/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 10/10/2018
Solution Control Lot: 16270 Bottle No.: 1076

Coordinator

Last Name: WATSON First Name: MATTHEW MI: R
Signature: *Matthew Watson* #2078 Badge No.: 7078
Date: 07/11/2017

*Black Key Temperature Probe Serial.....# DDXKP2-390 (MRW)
*Digital NIST Temperature Measuring System Serial.....# 150671438 (MRW)



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0051
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02636 Calib. Date: 07/11/2017 Calib. No.: 00033
Certification File No.: 02637 Cert. Date: 07/11/2017 Cert. No.: 00029
Linearity File No.: 02581 Lin. Date: 01/11/2017 Lin. No.: 00027
Solution File No.: 02635 Soln. Date: 06/01/2017 Soln. No.: 00273
Sequential File No.: 02637 File Date: 07/11/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339
Control Solution %: 0.100% Expires: 10/10/2018
Solution Control Lot: 16270 Bottle No.: 1076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:55D		
Control 1 EC	0.100%	10:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		
Control 2 EC	0.099%	10:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		
Control 3 EC	0.099%	10:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		

All tests within acceptable tolerance.

Coordinator

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature: Tpr. II Matthew A. Watson #7078

Badge No.: 7078

Date: 07/11/2017

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.