

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02175	Calib. Date: 01/11/2017
Certification File No.:	02176	Cert. No.: 00030
Linearity File No.:	02177	Cert. Date: 01/11/2017
Solution File No.:	02188	Lin. No.: 00026
Sequential File No.:	02188	Lin. Date: 01/11/2017
		Soln. No.: 00228
		Soln. Date: 02/13/2017
		File Date: 02/13/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0341
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0022

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:35S		
Control 1 EC	0.101%	08:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36S		
Control 2 EC	0.100%	08:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38S		
Control 3 EC	0.101%	08:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOLES

First Name: VITO

MI: M

Signature: _____

Badge No.: 171

Date: 02/13/2017

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02175 Calib. Date: 01/11/2017 Calib. No.: 00030
Certification File No.: 02176 Cert. Date: 01/11/2017 Cert. No.: 00026
Linearity File No.: 02177 Lin. Date: 01/11/2017 Lin. No.: 00026
Solution File No.: 02199 Soln. Date: 03/11/2017 Soln. No.: 00229
Sequential File No.: 02199 File Date: 03/11/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0137

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:19S		
Control 1 EC	0.100%	08:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20S		
Control 2 EC	0.099%	08:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22S		
Control 3 EC	0.100%	08:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMPSON

First Name: SHAUN

MI: .

Signature:  #142

Badge No.: 142

Date: 03/11/2017