

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02901	Cert. No.: 00033
Sequential File No.:	02901	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 07/24/2019
		Soln. No.: 00316
		File Date: 07/24/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1287

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:05D		
Control 1 EC	0.096%	00:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		
Control 2 EC	0.096%	00:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	00:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:07D		
Control 3 EC	0.096%	00:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_

Badge No.: 191

Date: 07/24/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02899	Cert. No.: 00033
Sequential File No.:	02899	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 07/19/2019
		Soln. No.: 00315
		File Date: 07/19/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1054

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:28D		
Control 1 EC	0.101%	15:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29D		
Control 2 EC	0.100%	15:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30D		
Control 3 EC	0.099%	15:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBONE

First Name: BRIAN

MI: M

Signature: \_\_\_\_\_

TPR. I B M #7029

Badge No.: 7029

Date: 07/19/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	02859	Calib. Date:	04/11/2019
Certification File No.:	02860	Cert. No.:	00037
Linearity File No.:	02861	Cert. Date:	04/11/2019
Solution File No.:	02893	Lin. No.:	00032
Sequential File No.:	02893	Soln. Date:	06/02/2019
		Soln. No.:	00314
		File Date:	06/02/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	1157

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:39D		
Control 1 EC	0.098%	02:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D		
Control 2 EC	0.097%	02:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	02:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:41D		
Control 3 EC	0.097%	02:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: \_\_\_\_\_

*PTL. Bristow #227*

Badge No.: 227

Date: 06/02/2019

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02859 Calib. Date: 04/11/2019 Calib. No.: 00037  
Certification File No.: 02860 Cert. Date: 04/11/2019 Cert. No.: 00033  
Linearity File No.: 02861 Lin. Date: 04/11/2019 Lin. No.: 00032  
Solution File No.: 02888 Soln. Date: 05/26/2019 Soln. No.: 00313  
Sequential File No.: 02888 File Date: 05/26/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339  
Control Solution %: 0.100% Expires: 09/17/2020  
Solution Control Lot: 18290 Bottle No.: 0931

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:36D		
Control 1 EC	0.097%	05:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:37D		
Control 2 EC	0.096%	05:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	05:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:39D		
Control 3 EC	0.096%	05:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: HURLEY

First Name: R

MI: J

Signature: \_\_\_\_\_

*PH. R. H. #205*

Badge No.: 205

Date: 05/26/2019



# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02882	Cert. No.: 00033
Sequential File No.:	02882	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 05/18/2019
		Soln. No.: 00312
		File Date: 05/18/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0959

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:02D		
Control 1 EC	0.098%	00:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:03D		
Control 2 EC	0.097%	00:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:04D		
Control 3 EC	0.097%	00:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK	First Name: JOSEPH	MI: -
Signature: <u></u>	Badge No.: 216	Date: 05/18/2019

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02859	Calib. Date: 04/11/2019
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/11/2019
Solution File No.:	02875	Cert. No.: 00033
Sequential File No.:	02875	Lin. Date: 04/11/2019
		Lin. No.: 00032
		Soln. Date: 05/04/2019
		Soln. No.: 00311
		File Date: 05/04/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1403

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:15D		
Control 1 EC	0.098%	03:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17D		
Control 2 EC	0.097%	03:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18D		
Control 3 EC	0.097%	03:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_

*Pr Joseph #216*

Badge No.: 216

Date: 05/04/2019



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02859 Calib. Date: 04/11/2019 Calib. No.: 00037  
Certification File No.: 02860 Cert. Date: 04/11/2019 Cert. No.: 00033  
Linearity File No.: 02861 Lin. Date: 04/11/2019 Lin. No.: 00032  
Solution File No.: 02865 Soln. Date: 04/13/2019 Soln. No.: 00310  
Sequential File No.: 02865 File Date: 04/13/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0339  
Control Solution %: 0.100% Expires: 09/17/2020  
Solution Control Lot: 18290 Bottle No.: 1399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:53D		
Control 1 EC	0.100%	01:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:54D		
Control 2 EC	0.097%	01:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:55D		
Control 3 EC	0.097%	01:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: 

Badge No.: 191

Date: 04/13/2019

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02853	Cert. No.: 00032
Sequential File No.:	02853	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 03/30/2019
		Soln. No.: 00308
		File Date: 03/30/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0935

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:35S		
Control 1 EC	0.099%	05:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36S		
Control 2 EC	0.097%	05:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:38S		
Control 3 EC	0.097%	05:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BRISTOW

First Name: MARK

MI: V

Signature: \_\_\_\_\_

PTL. Bristow #227

Badge No.: 227

Date: 03/30/2019



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02846	Cert. No.: 00032
Sequential File No.:	02846	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 03/20/2019
		Soln. No.: 00307
		File Date: 03/20/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1408

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:25S		
Control 1 EC	0.099%	02:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:26S		
Control 2 EC	0.097%	02:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27S		
Control 3 EC	0.098%	02:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_



Badge No.: 216

Date: 03/20/2019

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02841	Cert. No.: 00032
Sequential File No.:	02841	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 02/23/2019
		Soln. No.: 00306
		File Date: 02/23/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0308

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:24S		
Control 1 EC	0.100%	01:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:26S		
Control 2 EC	0.098%	01:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:27S		
Control 3 EC	0.098%	01:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: PENNINGTON

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 191

Date: 02/23/2019



# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02804	Calib. Date: 10/24/2018
Certification File No.:	02805	Calib. No.: 00036
Linearity File No.:	02806	Cert. Date: 10/24/2018
Solution File No.:	02835	Cert. No.: 00032
Sequential File No.:	02835	Lin. Date: 10/24/2018
		Lin. No.: 00031
		Soln. Date: 01/28/2019
		Soln. No.: 00305
		File Date: 01/28/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0493

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:13S		
Control 1 EC	0.100%	12:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15S		
Control 2 EC	0.098%	12:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16S		
Control 3 EC	0.098%	12:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: \_\_\_\_\_

*Sgt M Henkel 149*

Badge No.: 149

Date: 01/28/2019

