

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	00866	Calib. Date: 07/06/2010
Certification File No.:	00867	Calib. No.: 00011
Linearity File No.:	00868	Cert. Date: 07/06/2010
Solution File No.:	00935	Cert. No.: 00010
Sequential File No.:	00935	Lin. Date: 07/06/2010
		Lin. No.: 00010
		Soln. No.: 00093
		File Date: 12/11/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	10C077	Expires: 03/10/2012
		Bottle No.: 0570

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:59S		
Control 1 EC	0.103%	17:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:00S		
Control 2 EC	0.101%	18:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:02S		
Control 3 EC	0.100%	18:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: 

Badge No.: 149

Date: 12/11/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
<b>Location:</b>	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	00866	Calib. Date: 07/06/2010
Certification File No.:	00867	Cert. No.: 00011
Linearity File No.:	00868	Cert. Date: 07/06/2010
Solution File No.:	00922	Lin. No.: 00010
Sequential File No.:	00922	Lin. Date: 07/06/2010
		Soln. No.: 00092
		Soln. Date: 11/21/2010
		File Date: 11/21/2010
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
<b>Control Solution %:</b>	0.100%	Serial No.: DDUN S3-0338
<b>Solution Control Lot:</b>	10C077	Expires: 03/10/2012
		Bottle No.: 0568

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:31S		
Control 1 EC	0.102%	05:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	05:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:32S		
Control 2 EC	0.100%	05:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:34S		
Control 3 EC	0.099%	05:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	05:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: *Gerard Henkel 149*

Badge No.: 149

Date: 11/21/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	00866	Calib. Date: 07/06/2010
Certification File No.:	00867	Calib. No.: 00011
Linearity File No.:	00868	Cert. Date: 07/06/2010
Solution File No.:	00911	Cert. No.: 00010
Sequential File No.:	00911	Lin. Date: 07/06/2010
		Lin. No.: 00010
		Soln. Date: 10/29/2010
		Soln. No.: 00091
		File Date: 10/29/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	10C077	Expires: 03/10/2012
		Bottle No.: 0559

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:52D		
Control 1 EC	0.101%	16:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:54D		
Control 2 EC	0.101%	16:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55D		
Control 3 EC	0.099%	16:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CIACCIA

First Name: MIKE

MI: -

Signature: \_\_\_\_\_

Badge No.: 172

Date: 10/29/2010

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
 Location: PENNSAUKEN TWSP. P.D.  
 Calibration File No.: 00866 Calib. Date: 07/06/2010 Calib. No.: 00011  
 Certification File No.: 00867 Cert. Date: 07/06/2010 Cert. No.: 00010  
 Linearity File No.: 00868 Lin. Date: 07/06/2010 Lin. No.: 00010  
 Solution File No.: 00903 Soln. Date: 10/02/2010 Soln. No.: 00090  
 Sequential File No.: 00903 File Date: 10/02/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338  
 Control Solution %: 0.100% Expires: 03/10/2012  
 Solution Control Lot: 10C077 Bottle No.: 0555

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:05D		
Control 1 EC	0.100%	16:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06D		
Control 2 EC	0.099%	16:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08D		
Control 3 EC	0.099%	16:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KILLION

First Name: MICHAEL

MI: P.

Signature: \_\_\_\_\_

*Ptl. M. Killion #170*  
 Ptl. M. Killion #170

Badge No.: 170

Date: 10/02/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	Calib. No.:	00011
Calibration File No.:	00866	Cert. Date:	07/06/2010
Certification File No.:	00867	Lin. Date:	07/06/2010
Linearity File No.:	00868	Soln. Date:	09/04/2010
Solution File No.:	00890	File Date:	09/04/2010
Sequential File No.:	00890		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	091069	Expires:	09/18/2011
		Bottle No.:	0024

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:54D		
Control 1 EC	0.102%	15:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:55D		
Control 2 EC	0.100%	15:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57D		
Control 3 EC	0.099%	15:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: PH HENKEL 149

Badge No.: 149

Date: 09/04/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	00866	Calib. Date:	07/06/2010
Certification File No.:	00867	Cert. No.:	00011
Linearity File No.:	00868	Cert. Date:	07/06/2010
Solution File No.:	00882	Lin. No.:	00010
Sequential File No.:	00882	Soln. No.:	00088
		Soln. Date:	08/06/2010
		File Date:	08/06/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	09I069	Expires:	09/18/2011
		Bottle No.:	0018

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:31D		
Control 1 EC	0.101%	10:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33D		
Control 2 EC	0.100%	10:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34D		
Control 3 EC	0.100%	10:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SULZBACH

First Name: CHRIS

MI: L

Signature: \_\_\_\_\_



Badge No.: 127

Date: 08/06/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	00784	Calib. Date: 01/06/2010
Certification File No.:	00785	Calib. No.: 00010
Linearity File No.:	00786	Cert. Date: 01/06/2010
Solution File No.:	00856	Cert. No.: 00009
Sequential File No.:	00856	Lin. Date: 01/06/2010
		Lin. No.: 00009
		Soln. Date: 06/18/2010
		Soln. No.: 00086
		File Date: 06/18/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	09I069	Expires: 09/18/2011
		Bottle No.: 0022

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:10D		
Control 1 EC	0.101%	09:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		
Control 2 EC	0.099%	09:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13D		
Control 3 EC	0.099%	09:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SULZBACH

First Name: CHRIS

MI: L

Signature: \_\_\_\_\_



Badge No.: 127

Date: 06/18/2010

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	00784	Calib. Date: 01/06/2010
Certification File No.:	00785	Calib. No.: 00010
Linearity File No.:	00786	Cert. Date: 01/06/2010
Solution File No.:	00846	Cert. No.: 00009
Sequential File No.:	00846	Lin. Date: 01/06/2010
		Lin. No.: 00009
		Soln. Date: 05/19/2010
		Soln. No.: 00085
		File Date: 05/19/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	09I069	Expires: 09/18/2011
		Bottle No.: 0017

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:48D		
Control 1 EC	0.102%	16:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:49D		
Control 2 EC	0.100%	16:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:50D		
Control 3 EC	0.100%	16:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SULZBACH

First Name: CHRIS

MI: L

Signature: \_\_\_\_\_



Badge No.: 127

Date: 05/19/2010



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	Calib. No.:	00010
Calibration File No.:	00784	Cert. No.:	00009
Certification File No.:	00785	Lin. No.:	00009
Linearity File No.:	00786	Soln. No.:	00084
Solution File No.:	00830	Calib. Date:	01/06/2010
Sequential File No.:	00830	Cert. Date:	01/06/2010
		Lin. Date:	01/06/2010
		Soln. Date:	04/19/2010
		File Date:	04/19/2010
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0339
Solution Control Lot:	09I069	Expires:	09/18/2011
		Bottle No.:	0005

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:01D		
Control 1 EC	0.100%	00:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:03D		
Control 2 EC	0.098%	00:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:04D		
Control 3 EC	0.098%	00:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: *Gerald Henkel* 149

Badge No.: 149

Date: 04/19/2010