

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02684 Calib. Date: 11/27/2017 Calib. No.: 00034  
Certification File No.: 02685 Cert. Date: 11/27/2017 Cert. No.: 00030  
Linearity File No.: 02686 Lin. Date: 11/27/2017 Lin. No.: 00029  
Solution File No.: 02699 Soln. Date: 01/26/2018 Soln. No.: 00284  
Sequential File No.: 02699 File Date: 01/26/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338  
Control Solution %: 0.100% Expires: 10/17/2019  
Solution Control Lot: 17360 Bottle No.: 0955

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:26S		
Control 1 EC	0.099%	02:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28S		
Control 2 EC	0.097%	02:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:29S		
Control 3 EC	0.097%	02:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_



Badge No.: 216

Date: 01/26/2018