

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARUM-0066
Location: PENNSAUKEN TWSP. P.D.
Calibration File No.: 02340 Calib. Date: 11/27/2017 Calib. No.: 00032
Certification File No.: 02341 Cert. Date: 11/27/2017 Cert. No.: 00028
Linearity File No.: 02342 Lin. Date: 11/27/2017 Lin. No.: 00028
Solution File No.: 02359 Soln. Date: 01/26/2018 Soln. No.: 00244
Sequential File No.: 02359 File Date: 01/26/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0341
Control Solution %: 0.100% Expires: 10/17/2019
Solution Control Lot: 17360 Bottle No.: 0827

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:23S		
Control 1 EC	0.099%	02:24S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25S		
Control 2 EC	0.098%	02:25S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:26S		
Control 3 EC	0.099%	02:27S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: _____

Joe Kuchmek #216

Badge No.: 216

Date: 01/26/2018