



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02591	Cert. No.: 00028
Sequential File No.:	02591	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 02/06/2017
		Soln. No.: 00266
		File Date: 02/06/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0222

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.100%	07:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54S		
Control 2 EC	0.098%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 3 EC	0.099%	07:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 02/06/2017

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02602	Cert. No.: 00028
Sequential File No.:	02602	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 03/06/2017
		Soln. No.: 00267
		File Date: 03/06/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0357

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:07S		
Control 1 EC	0.099%	06:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:08S		
Control 2 EC	0.098%	06:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10S		
Control 3 EC	0.098%	06:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature:  \*187

Badge No.: 187

Date: 03/06/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017
Certification File No.:	02580	Calib. No.: 00032
Linearity File No.:	02581	Cert. Date: 01/11/2017
Solution File No.:	02608	Cert. No.: 00028
Sequential File No.:	02608	Lin. Date: 01/11/2017
		Lin. No.: 00027
		Soln. Date: 04/08/2017
		Soln. No.: 00268
		File Date: 04/08/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 2 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 3 EC	0.098%	09:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SAMPSON

First Name: OFFICER

MI: .

Signature: \_\_\_\_\_

*Thomas Sampson #142*

Badge No.: 142

Date: 04/08/2017