

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02524	Calib. Date: 08/22/2016
Certification File No.:	02525	Calib. No.: 00031
Linearity File No.:	02526	Cert. Date: 08/22/2016
Solution File No.:	02578	Cert. No.: 00027
Sequential File No.:	02578	Lin. Date: 08/22/2016
		Lin. No.: 00026
		Soln. Date: 01/07/2017
		Soln. No.: 00264
		File Date: 01/07/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0339
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0203

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:15S		
Control 1 EC	0.102%	00:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	00:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16S		
Control 2 EC	0.101%	00:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18S		
Control 3 EC	0.101%	00:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	00:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI	First Name: P	MI: E
Signature: <u> 0187</u>	Badge No.: 187	Date: 01/07/2017

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02579 Calib. Date: 01/11/2017 Calib. No.: 00032  
Certification File No.: 02580 Cert. Date: 01/11/2017 Cert. No.: 00028  
Linearity File No.: 02581 Lin. Date: 01/11/2017 Lin. No.: 00027  
Solution File No.: 02591 Soln. Date: 02/06/2017 Soln. No.: 00266  
Sequential File No.: 02591 File Date: 02/06/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338  
Control Solution %: 0.100% Expires: 10/28/2018  
Solution Control Lot: 16330 Bottle No.: 0222

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.100%	07:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54S		
Control 2 EC	0.098%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 3 EC	0.099%	07:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 02/06/2017

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0051  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 02579 Calib. Date: 01/11/2017 Calib. No.: 00032  
Certification File No.: 02580 Cert. Date: 01/11/2017 Cert. No.: 00028  
Linearity File No.: 02581 Lin. Date: 01/11/2017 Lin. No.: 00027  
Solution File No.: 02602 Soln. Date: 03/06/2017 Soln. No.: 00267  
Sequential File No.: 02602 File Date: 03/06/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338  
Control Solution %: 0.100% Expires: 10/28/2018  
Solution Control Lot: 16330 Bottle No.: 0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:07S		
Control 1 EC	0.099%	06:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:08S		
Control 2 EC	0.098%	06:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10S		
Control 3 EC	0.098%	06:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*[Handwritten Signature]* #187

Badge No.: 187

Date: 03/06/2017

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	02579	Calib. Date: 01/11/2017      Calib. No.: 00032
Certification File No.:	02580	Cert. Date: 01/11/2017      Cert. No.: 00028
Linearity File No.:	02581	Lin. Date: 01/11/2017      Lin. No.: 00027
Solution File No.:	02608	Soln. Date: 04/08/2017      Soln. No.: 00268
Sequential File No.:	02608	File Date: 04/08/2017
Calibrating Unit:	WET	Model No.: CU-34      Serial No.: DDUN S3-0338
Control Solution %:	0.100%	Expires: 10/28/2018
Solution Control Lot:	16330	Bottle No.: 0331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 2 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 3 EC	0.098%	09:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SAMPSON

First Name: OFFICER

MI: .

Signature: \_\_\_\_\_

*R. Sampson # 142*

Badge No.: 142

Date: 04/08/2017