

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Police Department 2400 Bethel Ave. Pennsauken, NJ 08109		2. Contact: Sgt. Chris Sulzbach	4. Date: 5/17/2016
		3. Phone Number: (856)488-0080 x2301	5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0066	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Complaint received of the instrument experiencing control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed .000% Electrochemical Sensor (EC) results on control tests.

Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 02114

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Stephen M Grusemeyer #6909
 Name & Badge Number (Print)

Stephen M Grusemeyer
 Signature

Date
5/17/2016