## STATE OF NEW JERSEY DEPARTMENT OF LAW AND PUBLIC SAFETY

BREATH TE	DIVISION OF STA STING INSTRUMEN		CE REPORT
1. Department:	2. Contact:		4. Date:
Pennsauken Police Department	Sgt. Cl	ris Sulzbach	5/17/2016
2400 Bethel Ave.	3. Phone N	umber:	5. County:
Pennsauken, NJ 08109	(856)48	88-0080 x2301	Camden
6. Alcotest Instrument Serial Number:	7. Simulator Component Seria	l Number:	8. Temperature Probe Component Serial Number:
ARUM-0066	N/A		N/A
9. Reason for Service:  Complaint received of the instrument experiencing control test failures.			
10. Comments:			
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed .000% Electrochemical Sensor (EC) results on control tests.			
Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.			
Last Known Sequential File #:02114			
11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.			
X 12. The above Instrument/Component is placed out of service pending further evaluation.			
13. The above Instrument/Component is placed back in service.			
"I have been certified by the Attorney Ger official capacity as a Breath Test Coord components of evidential breath test system capacity. The results of my inspection are accurate to the best of my knowledge. I a subject to punishment."	linator/Instructor I performs. I have inspected the erecorded on this form. I	m inspections of evidential breath test certify that the forg	t device listed on this report in my official going statements made by me are true and
Tpr. I Stephen M Grusemeyer #6	6909	Signature 5	Date 7/30X

S.P. 343 (Rev. 01/10) (S.O.P. F26)