

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 01740 Calib. Date: 12/09/2013 Calib. No.: 00023  
Certification File No.: 01741 Cert. Date: 12/09/2013 Cert. No.: 00020  
Linearity File No.: 01742 Lin. Date: 12/09/2013 Lin. No.: 00020  
Solution File No.: 01771 Soln. Date: 01/19/2014 Soln. No.: 00170  
Sequential File No.: 01771 File Date: 01/19/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 13I119 Bottle No.: 0382

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:43S		
Control 1 EC	0.100%	20:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44S		
Control 2 EC	0.100%	20:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46S		
Control 3 EC	0.101%	20:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_

*Pr JL #190*

Badge No.: 190

Date: 01/19/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01740	Calib. Date: 12/09/2013
Certification File No.:	01741	Calib. No.: 00023
Linearity File No.:	01742	Cert. Date: 12/09/2013
Solution File No.:	01783	Cert. No.: 00020
Sequential File No.:	01783	Lin. Date: 12/09/2013
		Lin. No.: 00020
		Soln. Date: 02/11/2014
		Soln. No.: 00171
		File Date: 02/11/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0376

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:19S		
Control 1 EC	0.102%	23:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:20S		
Control 2 EC	0.100%	23:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:22S		
Control 3 EC	0.100%	23:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_

P. J. # 190

Badge No.: 190

Date: 02/11/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01740	Calib. Date: 12/09/2013
Certification File No.:	01741	Calib. No.: 00023
Linearity File No.:	01742	Cert. Date: 12/09/2013
Solution File No.:	01793	Cert. No.: 00020
Sequential File No.:	01793	Lin. Date: 12/09/2013
		Lin. No.: 00020
		Soln. Date: 03/11/2014
		Soln. No.: 00172
		File Date: 03/11/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0383

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:03S		
Control 1 EC	0.101%	16:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04S		
Control 2 EC	0.100%	16:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06S		
Control 3 EC	0.101%	16:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature:  #190

Badge No.: 190

Date: 03/11/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	01740	Calib. Date:	12/09/2013
Certification File No.:	01741	Cert. Date:	12/09/2013
Linearity File No.:	01742	Lin. Date:	12/09/2013
Solution File No.:	01804	Soln. Date:	04/04/2014
Sequential File No.:	01804	File Date:	04/04/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0377

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:43S		
Control 1 EC	0.101%	23:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:45S		
Control 2 EC	0.099%	23:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46S		
Control 3 EC	0.099%	23:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_



Badge No.: 187

Date: 04/04/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	01740	Calib. Date:	12/09/2013
Certification File No.:	01741	Cert. No.:	00023
Linearity File No.:	01742	Cert. Date:	12/09/2013
Solution File No.:	01819	Lin. No.:	00020
Sequential File No.:	01819	Soln. Date:	04/19/2014
		Soln. No.:	00174
		File Date:	04/19/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	13I119	Expires:	09/10/2015
		Bottle No.:	0391

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:30D		
Control 1 EC	0.103%	04:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	04:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:31D		
Control 2 EC	0.102%	04:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	04:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:33D		
Control 3 EC	0.101%	04:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_

Badge No.: 190

Date: 04/19/2014

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 01740 Calib. Date: 12/09/2013 Calib. No.: 00023  
Certification File No.: 01741 Cert. Date: 12/09/2013 Cert. No.: 00020  
Linearity File No.: 01742 Lin. Date: 12/09/2013 Lin. No.: 00020  
Solution File No.: 01823 Soln. Date: 04/30/2014 Soln. No.: 00175  
Sequential File No.: 01823 File Date: 04/30/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0338  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 13I119 Bottle No.: 0396

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:33D		
Control 1 EC	0.101%	15:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35D		
Control 2 EC	0.100%	15:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:36D		
Control 3 EC	0.100%	15:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERN

First Name: RAYMOND

MI: T

Signature: \_\_\_\_\_

*R Kern #181*

Badge No.: 181

Date: 04/30/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01740	Calib. Date: 12/09/2013
Certification File No.:	01741	Cert. No.: 00023
Linearity File No.:	01742	Cert. Date: 12/09/2013
Solution File No.:	01830	Lin. No.: 00020
Sequential File No.:	01830	Lin. Date: 12/09/2013
		Soln. No.: 00176
		Soln. Date: 05/23/2014
		File Date: 05/23/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0338
Solution Control Lot:	13I119	Expires: 09/10/2015
		Bottle No.: 0397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:06D		
Control 1 EC	0.099%	20:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:07D		
Control 2 EC	0.098%	20:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:08D		
Control 3 EC	0.099%	20:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_

*PR JL #190*

Badge No.: 190

Date: 05/23/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	01740	Calib. Date:	12/09/2013
Certification File No.:	01741	Cert. No.:	00023
Linearity File No.:	01742	Lin. Date:	12/09/2013
Solution File No.:	01841	Cert. No.:	00020
Sequential File No.:	01841	Lin. No.:	00020
		Soln. Date:	06/07/2014
		Soln. No.:	00177
		File Date:	06/07/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0338
Solution Control Lot:	131119	Expires:	09/10/2015
		Bottle No.:	0393

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:48D		
Control 1 EC	0.097%	06:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:49D		
Control 2 EC	0.097%	06:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:50D		
Control 3 EC	0.098%	06:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KUCHMEK

First Name: JOSEPH

MI: -

Signature: \_\_\_\_\_

*Joseph #190*

Badge No.: 190

Date: 06/07/2014



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 01843 Calib. Date: 06/09/2014 Calib. No.: 00024  
Certification File No.: 01844 Cert. Date: 06/09/2014 Cert. No.: 00021  
Linearity File No.: 01845 Lin. Date: 06/09/2014 Lin. No.: 00021  
Solution File No.: 01857 Soln. Date: 07/07/2014 Soln. No.: 00179  
Sequential File No.: 01857 File Date: 07/07/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 131119 Bottle No.: 0380

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:31D		
Control 1 EC	0.101%	22:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:32D		
Control 2 EC	0.100%	22:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34D		
Control 3 EC	0.100%	22:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*FABIANI* 187

Badge No.: 187

Date: 07/07/2014

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUM-0066  
Location: PENNSAUKEN TWSP. P.D.  
Calibration File No.: 01843 Calib. Date: 06/09/2014 Calib. No.: 00024  
Certification File No.: 01844 Cert. Date: 06/09/2014 Cert. No.: 00021  
Linearity File No.: 01845 Lin. Date: 06/09/2014 Lin. No.: 00021  
Solution File No.: 01866 Soln. Date: 08/02/2014 Soln. No.: 00180  
Sequential File No.: 01866 File Date: 08/02/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0340  
Control Solution %: 0.100% Expires: 09/10/2015  
Solution Control Lot: 13I119 Bottle No.: 0385

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:05D		
Control 1 EC	0.101%	18:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		
Control 2 EC	0.100%	18:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07D		
Control 3 EC	0.101%	18:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: HENKEL

First Name: GERALD

MI: F.

Signature: Sgt M Henkel 149

Badge No.: 149

Date: 08/02/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	01843	Calib. Date:	06/09/2014
Certification File No.:	01844	Cert. Date:	06/09/2014
Linearity File No.:	01845	Lin. Date:	06/09/2014
Solution File No.:	01877	Soln. Date:	09/01/2014
Sequential File No.:	01877	File Date:	09/01/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	131119	Expires:	09/10/2015
		Bottle No.:	0379

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:15D		
Control 1 EC	0.101%	17:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:16D		
Control 2 EC	0.100%	17:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 3 EC	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*FABIANI 187*

Badge No.: 187

Date: 09/01/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.		
Calibration File No.:	01843	Calib. Date:	06/09/2014
Certification File No.:	01844	Cert. No.:	00024
Linearity File No.:	01845	Cert. Date:	06/09/2014
Solution File No.:	01885	Lin. No.:	00021
Sequential File No.:	01885	Soln. Date:	09/29/2014
		Soln. No.:	00182
		File Date:	09/29/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0340
Solution Control Lot:	14F130	Expires:	06/20/2016
		Bottle No.:	1312

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:46D		
Control 1 EC	0.102%	20:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	20:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:47D		
Control 2 EC	0.102%	20:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:48D		
Control 3 EC	0.101%	20:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*[Handwritten Signature]*

Badge No.: 187

Date: 09/29/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01843	Calib. Date: 06/09/2014
Certification File No.:	01844	Calib. No.: 00024
Linearity File No.:	01845	Cert. Date: 06/09/2014
Solution File No.:	01894	Cert. No.: 00021
Sequential File No.:	01894	Lin. Date: 06/09/2014
		Lin. No.: 00021
		Soln. Date: 10/27/2014
		Soln. No.: 00183
		File Date: 10/27/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	14F130	Expires: 06/20/2016
		Bottle No.: 1306

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:27S		
Control 1 EC	0.102%	21:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:29S		
Control 2 EC	0.101%	21:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:30S		
Control 3 EC	0.101%	21:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*[Handwritten Signature]* 187

Badge No.: 187

Date: 10/27/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARUM-0066
Location:	PENNSAUKEN TWSP. P.D.	
Calibration File No.:	01843	Calib. Date: 06/09/2014
Certification File No.:	01844	Calib. No.: 00024
Linearity File No.:	01845	Cert. Date: 06/09/2014
Solution File No.:	01905	Cert. No.: 00021
Sequential File No.:	01905	Lin. Date: 06/09/2014
		Lin. No.: 00021
		Soln. Date: 11/25/2014
		Soln. No.: 00184
		File Date: 11/25/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUN S3-0340
Solution Control Lot:	14F130	Expires: 06/20/2016
		Bottle No.: 1307

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:04S		
Control 1 EC	0.103%	19:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	19:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:05S		
Control 2 EC	0.101%	19:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:07S		
Control 3 EC	0.101%	19:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: FABIANI

First Name: P

MI: E

Signature: \_\_\_\_\_

*[Handwritten Signature]* #187

Badge No.: 187

Date: 11/25/2014